



Gleichwertigkeitsnachweis für die Messergebnisse der Partikelmonitore mit dem Referenzverfahren - PM10, 2012



GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
Starting month:	Spring	Summer	Fall	Winter	
	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	$0,772y + 1,533$		N (Spring)	0	n
Regression (=0)	$0,812y$		N (Summer)	0	n
N	361	n	N (Fall)	0	n
			N (Winter)	0	n
Outliers	13	n	Outliers	9	%
Outliers	4%	%	Outliers	2%	%
Mean CM	25,7	µg/m³	Mean CM	20,9	µg/m³
Mean RM	21,4	µg/m³	Mean RM	21,4	µg/m³
Number of RM > 0.5LV	99	n	Number of CM > 0.5LV	96	n
Number of RM > LV	16	n	Number of CM > LV	21	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1,296	significant	Slope b	1,047	significant
Uncertainty of b	0,015		Uncertainty of b	0,012	
Intercept a	-1,986	significant	Intercept a	-1,497	significant
Uncertainty of a	0,386		Uncertainty of a	0,313	
r²	0,951		r²	0,951	
Slope b forced trough origin	1,231	significant			
Uncertainty of b (forced)	0,0084				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,85	µg/m³	Calibration	$0,812y + 0$	
Uncertainty of calibration (forced)	0,42	µg/m³	u(calibration)	0,42	µg/m³
Random term	4,02	µg/m³	Random term	3,26	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	12,79	µg/m³	Bias at LV	0,85	µg/m³
Combined uncertainty	13,41	µg/m³	Combined uncertainty	3,37	µg/m³
Expanded relative uncertainty	53,6%	fail	Expanded relative uncertainty	13,5%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss		u(b)
90182	119067	71832	5962		0,015
Raw data, slope forced trough origin					
Sxy	Syy	Sxx		u(b)[MaxLike] =0	u(b_forced)
288505	357587	236731		0,008	0,008
Calibrated, free intercept					
dxdy	dyy	dxx	rss		u(b)
73250	78554	71832	3910		0,012
BACKGROUND AUTOMATION					
Chart descriptions		Conf. Lvl List	Calibration List	Sdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%		0	0,00838
Description y-axis	CM PM10 (µg/m³)	97,5%	Free regression		
Conf. Lvl	Calibration Type	Filter List	Trough origin	Calib. In use (a)	Calib. In use (b)
2,25	-1	Exclude		0,000	0,812
Calibration uncertainty	Calibration_a	Calibration_b	u(bs_reference)	Cl Regression	Cl Calibrated
0,42	0,000	0,812	0,67	9,17	9,22



PM10_EDM-Cottbus

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijik (ruben.beijik@rivm.nl)



GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
	Spring	Summer	Fall	Winter	
Starting month:	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0,997y + -0,169		N (Spring)	0	n
Regression (≠0)	0,992y		N (Summer)	0	n
N	341	n	N (Fall)	0	n
			N (Winter)	0	n
Outliers	14	n	Outliers	12	%
Outliers	4%	%	Outliers	4%	%
Mean CM	22,7	µg/m³	Mean CM	22,5	µg/m³
Mean RM	22,5	µg/m³	Mean RM	22,5	µg/m³
Number of RM > 0.5LV	102	n	Number of CM > 0.5LV	106	n
Number of RM > LV	19	n	Number of CM > LV	19	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1,003		Slope b	0,995	
Uncertainty of b	0,011		Uncertainty of b	0,011	
Intercept a	0,170		Intercept a	0,172	
Uncertainty of a	0,290		Uncertainty of a	0,287	
r²	0,961		r²	0,961	
Slope b forced trough origin	1,008	significant			
Uncertainty of b (forced)	0,0059				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	0,61	µg/m³	Calibration	0,992y + 0	
Uncertainty of calibration (forced)	0,30	µg/m³	u(calibration)	0,30	µg/m³
Random term	2,85	µg/m³	Random term	2,84	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	0,31	µg/m³	Bias at LV	-0,10	µg/m³
Combined uncertainty	2,87	µg/m³	Combined uncertainty	2,85	µg/m³
Expanded relative uncertainty	11,5%	pass	Expanded relative uncertainty	11,4%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss	u(b)	
71752	73399	73010	2912	0,011	
Raw data, slope forced trough origin					
Sxy	Syy	Sxx	u(b)[MaxLike] ≠0		u(b_forced)
246050	249486	245537	0,006		0,006
Calibrated, free intercept					
dxdy	dyy	dxx	rss	u(b)	
71178	72230	73010	2865	0,011	
BACKGROUND AUTOMATION					
Chart descriptions		Conf. Lvl List	Calibration List	Sdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%	Free regression Trough origin	0	0,00592
Description y-axis	CM PM10 (µg/m³)	97,5%		Calib. In use (a)	Calib. In use (b)
Conf. Lvl	Calibration Type	Filter List		0,000	0,992
2,25	-1	Exclude		CI Regression	CI Calibrated
Calibration uncertainty	Calibration_a	Calibration_b	u(bs_reference)	6,60	6,63
0,30	0,000	0,992			



PM10-Sharp Cottbus

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijik (ruben.beijik@rivm.nl)



GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
	Spring	Summer	Fall	Winter	
Starting month:	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0,924y + 0,277		N (Spring)	0	n
Regression (≠0)	0,933y		N (Summer)	0	n
N	182		N (Fall)	0	n
			N (Winter)	0	n
Outliers	8	n	Outliers	4	%
Outliers	4%	%	Outliers	2%	%
Mean CM	22,4	µg/m³	Mean CM	20,9	µg/m³
Mean RM	20,9	µg/m³	Mean RM	20,9	µg/m³
Number of RM > 0.5LV	39	n	Number of CM > 0.5LV	42	n
Number of RM > LV	6	n	Number of CM > LV	7	n
REGRESSION RESULTS (RAW)			REGRESSION RESULTS (CALIBRATED)		
Slope b	1,083	significant	Slope b	1,008	
Uncertainty of b	0,020		Uncertainty of b	0,019	
Intercept a	-0,300		Intercept a	-0,230	
Uncertainty of a	0,491		Uncertainty of a	0,458	
r²	0,935		r²	0,935	
Slope b forced trough origin	1,072	significant			
Uncertainty of b (forced)	0,0099				
EQUIVALENCE TEST (RAW)			EQUIVALENCE TEST (CALIBRATED)		
Uncertainty of calibration	1,13	µg/m³	Calibration	0,933y + 0	
Uncertainty of calibration (forced)	0,50	µg/m³	u(calibration)	0,50	µg/m³
Random term	3,18	µg/m³	Random term	2,99	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	3,84	µg/m³	Bias at LV	0,16	µg/m³
Combined uncertainty	4,98	µg/m³	Combined uncertainty	3,00	µg/m³
Expanded relative uncertainty	19,9%	pass	Expanded relative uncertainty	12,0%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss	u(b)	
25802	28811	24704	1899	0,020	
Raw data, slope forced trough origin					
Sxy	Syy	Sxx	u(b)[MaxLike] ≠0		u(b_forced)
111143	119994	104578	0,010		0,010
Calibrated, free intercept					
dxdy	dyy	dxx	rss	u(b)	
24075	25083	24704	1649	0,019	
BACKGROUND AUTOMATION					
Chart descriptions		Confi. Lvl List	Calibration List	Sdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%	Free regression Trough origin	0	0,00995
Description y-axis	CM PM10 (µg/m³)	97,5%		Calib. In use (a)	Calib. In use (b)
Confli. Lvl	Calibration Type	Filter List		0,000	0,933
2,26	-1	Exclude		CI Regression	CI Calibrated
Calibration uncertainty	Calibration_a	Calibration_b	u(bs_reference)	7,34	7,43
0,50	0,000	0,933	0,67		



PM10-EDM-Eberswalde-Verkehr

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijk (ruben.beijk@rivm.nl)



GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
	Spring	Summer	Fall	Winter	
Starting month:	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0,807y + 1,349		N (Spring)	0	n
Regression (≠0)	0,847y		N (Summer)	0	n
N	338	n	N (Fall)	0	n
			N (Winter)	0	n
Outliers	11	n	Outliers	2	%
Outliers	3%	%	Outliers	1%	%
Mean CM	22,7	µg/m³	Mean CM	19,3	µg/m³
Mean RM	19,7	µg/m³	Mean RM	19,7	µg/m³
Number of RM > 0.5LV	74	n	Number of CM > 0.5LV	76	n
Number of RM > LV	14	n	Number of CM > LV	16	n
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)			
Slope b	1,240	significant	Slope b	1,044	significant
Uncertainty of b	0,017		Uncertainty of b	0,014	
Intercept a	-1,673	significant	Intercept a	-1,303	significant
Uncertainty of a	0,402		Uncertainty of a	0,341	
r²	0,936		r²	0,936	
Slope b forced trough origin	1,181	significant			
Uncertainty of b (forced)	0,0094				
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)			
Uncertainty of calibration	0,94	µg/m³	Calibration	0,847y + 0	
Uncertainty of calibration (forced)	0,47	µg/m³	u(calibration)	0,47	µg/m³
Random term	4,07	µg/m³	Random term	3,45	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	10,32	µg/m³	Bias at LV	0,91	µg/m³
Combined uncertainty	11,09	µg/m³	Combined uncertainty	3,57	µg/m³
Expanded relative uncertainty	44,4%	fail	Expanded relative uncertainty	14,3%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss	u(b)	
68125	86718	57200	5720	0,017	
Raw data, slope forced trough origin					
Sxy	Syy	Sxx	u(b)[MaxLike] ≠0		u(b_forced)
219584	261640	188344	0,010		0,009
Calibrated, free intercept					
dxdy	dyy	dxx	rss	u(b)	
57697	62203	57200	4077	0,014	
BACKGROUND AUTOMATION					
Chart descriptions		Conf. Lvl List	Calibration List	Stdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%	Free regression Trough origin	0	0,00942
Description y-axis	CM PM10 (µg/m³)	97,5%		Calib. In use (a)	Calib. In use (b)
Conf. Lvl	Calibration Type	Filter List		0,000	0,847
2,25	-1	Exclude			
Calibration uncertainty	Calibration_a	Calibration_b	u(bs_reference)	CI Regression	CI Calibrated
0,47	0,000	0,847	0,67	9,29	9,35



PM10-EDM-Eisenhüttenstadt

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijik (ruben.beijik@rivm.nl)



GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
	Spring	Summer	Fall	Winter	
Starting month:	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	1,248y + -1,603		N (Spring)	0	n
Regression (≠0)	1,189y		N (Summer)	0	n
N	335	n	N (Fall)	0	n
			N (Winter)	0	n
Outliers	10	n	Outliers	18	%
Outliers	3%	%	Outliers	5%	%
Mean CM	21,9	µg/m³	Mean CM	26,0	µg/m³
Mean RM	25,7	µg/m³	Mean RM	25,7	µg/m³
Number of RM > 0.5LV	124	n	Number of CM > 0.5LV	137	n
Number of RM > LV	26	n	Number of CM > LV	25	n
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)			
Slope b	0,802	significant	Slope b	0,960	significant
Uncertainty of b	0,012		Uncertainty of b	0,015	
Intercept a	1,285	significant	Intercept a	1,357	significant
Uncertainty of a	0,358		Uncertainty of a	0,426	
r²	0,923		r²	0,923	
Slope b forced trough origin	0,841	significant			
Uncertainty of b (forced)	0,0060				
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)			
Uncertainty of calibration	0,71	µg/m³	Calibration	1,189y + 0	
Uncertainty of calibration (forced)	0,30	µg/m³	u(calibration)	0,30	µg/m³
Random term	3,04	µg/m³	Random term	3,66	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	-8,64	µg/m³	Bias at LV	-0,65	µg/m³
Combined uncertainty	9,16	µg/m³	Combined uncertainty	3,72	µg/m³
Expanded relative uncertainty	36,6%	fail	Expanded relative uncertainty	14,9%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss	u(b)	
48954	41202	63040	3225	0,012	
Raw data, slope forced trough origin					
Sxy	Syy	Sxx	u(b)[Max_Like] ≠0		u(b_forced)
236764	201140	283577	0,006		0,006
Calibrated, free intercept					
dxdy	dyy	dxx	rss	u(b)	
58213	58261	63040	4588	0,015	
BACKGROUND AUTOMATION					
Chart descriptions		Conf. Lvl List	Calibration List	Stdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%	Free regression Trough origin	0	0,00605
Description y-axis	CM PM10 (µg/m³)	97,5%		Calib. In use (a)	Calib. In use (b)
Conf. Lvl	Calibration Type	Filter List		0,000	1,189
2,25	-1	Exclude		CI Regression	CI Calibrated
Calibration uncertainty	Calibration .a	Calibration .b	u(bs_reference)	7,01	7,04
0,30	0,000	1,189			



PM10_TEOM-SES Frankfurt Verkehr

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijik (ruben.beijik@rivm.nl)

GENERAL SETTINGS					
Substance	Unit	Limit value	RM uncertainty	Confidence Level	Max Uncertainty
PM10	µg/m³	50	0,67	0,975	25%
	Spring	Summer	Fall	Winter	
Starting month:	3	6	9	12	
DATA SELECTION					
	Column	Value	Exclude instead of exclusive?		Status
Filter 1					Ignore
Filter 2					Ignore
Filter 3					Ignore
Filter 4					Ignore
CALIBRATION SETTING					
Calibration based on:	SLOPE TROUGH ORIGIN		OK		
Comments:					
RAW DATA			RESULTS AFTER CALIBRATING		
Regression	0,76y + 1,776		N (Spring)	0	n
Regression (≠0)	0,819y		N (Summer)	0	n
N	344	n	N (Fall)	0	n
			N (Winter)	0	n
Outliers	11	n	Outliers	5	%
Outliers	3%	%	Outliers	1%	%
Mean CM	20,2	µg/m³	Mean CM	16,6	µg/m³
Mean RM	17,1	µg/m³	Mean RM	17,1	µg/m³
Number of RM > 0.5LV	54	n	Number of CM > 0.5LV	60	n
Number of RM > LV	7	n	Number of CM > LV	8	n
REGRESSION RESULTS (RAW)		REGRESSION RESULTS (CALIBRATED)			
Slope b	1,316	significant	Slope b	1,070	significant
Uncertainty of b	0,019		Uncertainty of b	0,015	
Intercept a	-2,337	significant	Intercept a	-1,782	significant
Uncertainty of a	0,383		Uncertainty of a	0,314	
r²	0,929		r²	0,929	
Slope b forced trough origin	1,220	significant			
Uncertainty of b (forced)	0,0104				
EQUIVALENCE TEST (RAW)		EQUIVALENCE TEST (CALIBRATED)			
Uncertainty of calibration	1,01	µg/m³	Calibration	0,819y + 0	
Uncertainty of calibration (forced)	0,52	µg/m³	u(calibration)	0,52	µg/m³
Random term	3,87	µg/m³	Random term	3,18	µg/m³
Additional uncertainty (optional)	0,00	µg/m³	Additional uncertainty (optional)	0,00	µg/m³
Bias at LV	13,45	µg/m³	Bias at LV	1,74	µg/m³
Combined uncertainty	13,99	µg/m³	Combined uncertainty	3,62	µg/m³
Expanded relative uncertainty	56,0%	fail	Expanded relative uncertainty	14,5%	pass
Ref sampler uncertainty	0,67	µg/m³	Ref sampler uncertainty	0,67	µg/m³
Limit value	50	µg/m³	Limit value	50	µg/m³
STATISTICAL INFORMATION					
Raw data, free intercept					
dxdy	dyy	dxx	rss		u(b)
53428	72226	42539	5272		0,019
Raw data, slope forced trough origin					
Sxy	Syy	Sxx		u(b)[MaxLike] ≠0	u(b_forced)
172605	212778	143592		0,010	0,010
Calibrated, free intercept					
dxdy	dyy	dxx	rss		u(b)
43783	48502	42539	3511		0,015
BACKGROUND AUTOMATION					
Chart descriptions		Conf. Lvl List	Calibration List	Sdev of all calibrations in use	
Description x-axis	RM PM10 (µg/m³)	97,5%		0	0,01037
Description y-axis	CM PM10 (µg/m³)	97,5%	Free regression		
Conf. Lvl	Calibration Type	Filter List	Trough origin	Calib. In use (a)	Calib. In use (b)
2,25	-1	Exclude		0,000	0,819
Calibration uncertainty	Calibration_a	Calibration_b	u(bs_reference)	CI Regression	CI Calibrated
0,52	0,000	0,819	0,67	8,84	8,92



PM10 EDM-Hasenholz

© Copyright RIVM 2011, The Netherlands
Version 2.9
Author: Ruben Beijik (ruben.beijik@rivm.nl)